Section IV: 510k Summary

Applicant's Identification

Applicant

Grandway Technology (Shenzhen) Limited

Phone Number

(00852)-2851-6789

Fax Number

(00852)-2851-6278

Contact Person

Mr. Patrick Chow

Date of Application

31st Jan, 2011

Device's Identification

Device Proprietary Name

Digital Automatic Wrist Blood Pressure Monitor WBPM10

Series

Model No.: MD10xy

x --- The first character (0-0 or A-Z) is for the identification of cabinet (housing).

y --- The second character (0-9 or A-Z) is for the minor change revision of device. The mentioned "minor change" refers to those device changes not to be affected the conformity test results of EMC & safety as well as device performance, i.e. IEC 60601-1 and EN 60601-1-2.

Common Name
Classification Name

Non-invasive Blood Pressure Measurement System Non-invasive Blood Pressure Measurement System

(Class II per 21 CFR 870.1130)

Marketed Devices to which Equivalence is Claimed

DEVICE	MANUFACTURER	510(k) Number
AViTA BPM1 Series (Wrist Type) Blood	AViTA Corp	K030230
Pressure Meter (or Monitor)		

Device Description

Digital Automatic Wrist Blood Pressure Monitor WBPM10 Series is a non-invasive blood pressure measurement system for use by medical professional or at home. It is designed to measure the systolic and diastolic blood pressure, and pulse rate (heartbeat rate) of an individual in each measurement and then displays the readings on a digital panel.

The WBPM10 Series utilizes the oscillometric methodology, in which an inflatable cuff is $\mathbb{P}^{2/3}$ wrapped around the wrist of an individual, for the blood pressure measurement. This means the monitor detects your blood's movement through your brachial artery and converts the movements into a digital reading.

Intended Use (Indication for Use)

Digital Automatic Wrist Blood Pressure Monitor WBPM10 Series is for use by medical professional or at home. The WBPM10 Series is intended to measure the systolic and diastolic blood pressure, and pulse rate (heartbeat rate) of an individual by using a non-invasive technique, in which an inflatable cuff is wrapped around the wrist of an individual.

Comparison of Technological Characteristics between New Device and Predicate Devices

The Digital Automatic Wrist Blood Pressure Monitor WBPM10 Series is a non-invasive measuring device and utilizes the oscilliometric methodology to measure the blood pressure reading. The key components of device are a pressure sensor, electric valve and an electronic control module together with an electric pump, which inflate (and deflate) the inflatable cuff automatically according to our designed architecture. The predicate device adopts exactly same methodology and key components for measuring the blood pressure reading.

Clinical & Non-clinical Tests

A systematic & independent clinical test was conducted to validate the performance of the Digital Automatic Wrist Blood Pressure Monitor WBPM10 Series. The results demonstrated that WBPM10 Series meets the requirement of ANSI/AAMI SP-10-2002.

Comprehensive safety and EMC tests were performed and compiled to demonstrate WBPM10 Series is safe for use. Tests include

EN1060-1:1995/A2:2009

EN1060-3:1997/A2:2009

IEC60601-1:2005 + CORR. 1 (2006) + CORR. 2 (2007)

EN60601-1-2:2007

FCC Part 15

ISO10993-5:2009

ISO10993-10:2002 + A1:2006

EN60601-1-4:2007

Conclusion

Digital Automatic Wrist Blood Pressure Monitor WBPM10 Series has the same intended use and similar technological characteristics as predicate device (K030230). Moreover,

K110299 + P3/3

bench testing contained in this submission and clinical testing supplied demonstrated that no differences in the technological characteristics and questioning on safety or effectiveness to be raised. Thus, the Digital Automatic Wrist Blood Pressure Monitor WBPM10 Series is substantially equivalent to the predicate device.





Food and Drug Administration 10903 New Hampshire Avenue Document Control Room –WO66-G609 Silver Spring, MD 20993-0002

Grandway Technology (Shenzhen) Limited c/o Mr. Patrick Chow General Manager Building 6 and 7 Zhu Keng Industrial Zone, Ping Shan, Long Gang District Shenshen, Guang Dong China 518118

MAR 2 1 2011

Re: K110299

Trade/Device Name: Digital Automatic Wrist Blood Pressure Monitor WBPM10 Series

Regulatory Number: 21 CFR 870.1130

Regulation Name: Non-invasive Blood Pressure Measurement System

Regulatory Class: II (two) Product Code: DXN Dated: January 26, 2011 Received: February 1, 2011

Dear Mr. Chow:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the <u>Federal Register</u>.

Page 2 - Mr. Patrick Chow

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm.

Sincerely yours,

Bram D. Zuckerman, M.D.

Director

Division of Cardiovascular Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

510(k) Number (if known):		
Device Name:		
Digital Automatic Wrist Blood Pr	ressure Monito	or WBPM10 Series
	Model No.:	MD10xy
y The second character (0-9 mentioned "minor change" re	or A-Z) is for efers to those	dentification of cabinet (housing). The minor change revision of device. The device changes not to be affected the sell as device performance, i.e. IEC 60601-1
Indications For Use:		
professional or at home. The V	WBPM10 Serie pulse rate (he	itor WBPM10 Series is for use by medicales is intended to measure the systolic and eartbeat rate) of an individual by using a le cuff is wrapped around the wrist of an
Prescription Use (Part 21 CFR 801 Subpart D)	AND/OR	Over-The-Counter UseX(21 CFR 801 Subpart C)
(PLEASE DO NOT WRITE BELOW	/ THIS LINE-CO	NTINUE ON ANOTHER PAGE IF NEEDED)
Concurrence of CDRH,	<i>T</i> V //	
(Division Signature) Big(k) Num		Packer man 3/21/2011 ar Devices 0299